

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

DEC 29 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130132</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Matthew E Armstrong</u>	ID# <u>5906</u> Date <u>December 20, 2010</u>
A Agency <u>Seward Police Department</u> Phone # <u>(907) 224-3338</u>	
Instrument Location <u>410 Adams St Seward Community Jail, Seward AK</u>	
B Alco S/N <u>G30459</u> Target Value <u>.078</u> High Pressure <u>300</u>	
Alco Test Values <u>.076</u> <u>.076</u> 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u> <u>COB</u> <u>12/30/10</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 9<sup>th</sup> day of MAR, 2011

[Signature]  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130132  
003 12/29/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

DECEMBER 28, 2010 003 12/30/10

OPERATOR'S NAME:  
ARMSTRONG, MATTHEW E  
OPERATOR'S NUMBER: 5906  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
TEST

O.L. #: 0000000  
DEPT/AGENCY: SMD1  
CASE/REPORT: 10-0000  
TEST TYPE: U  
ALCO TARGET VALUE: .078  
ALCO S/N: 630459

### — BREATH ANALYSIS —

.078 ADJUSTED FOR 29.52 in  
ALCO TARGET .076 13:04  
BLANK TEST .000 13:05  
INTERNAL STANDARD VERIFIED 13:05  
ALCO TO 29.52 in .076 13:05  
BLANK TEST .000 13:06  
SUBJECT SAMPLE .000 13:06  
BLANK TEST .000 13:07  
ALCO TO 29.54 in .076 13:07  
BLANK TEST .000 13:08

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130132

DECEMBER 28, 2010  
TIME 13:00

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 03/29/00  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 45c  
BAROMETER: 29.52 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~